

## Certificate of Analysis

Product name	CCA-030 MC Reagent Calcium Assay Kit (OCPC)
--------------	---

Lot Number : 6C9301

### Performance

	Guaranteed value
Measurement range	0.2~30 mg/dL
$\Delta OD_{570nm}$	0.2~0.6
C.V.(%)	< 5 %

### Accuracy (measuring serum)

	calcium concentration (mg/dL)
ICP certificated	6.9±1.0
With the kit	6.7±0.2